sales@murada.co.id
www.murada.co.id



REED INSTRUMENTS Humidity Calibration Standard 33% RH, 161C50

Item **32LVJ9331** Mfr. Model **R9933**

Product Details

Item Reference: **161C50**	Width (in): **2**
Brand: **REED INSTRUMENTS**	Height (in): **6.5**
Weight (lb): **0.75**	Harmonization Code: **9031900000**
Length (in): **5.25**	

Product Description

Humidity Calibration Standard Relative Humidity Accuracy +/-1 Percent at 77 Degrees F Number of Pieces 1 Compatible Probe Diameter Up to 0.6 in Compatible Manufacturer Model Number 8706 R6001 R6200 R9900 SD-3007 Product Type Humidity Meter Calibration Standard Sample Humidity 33 Percent Function Verification and Calibration of Humidity Probes